



Spectrum processing :
No peaks omitted

Processing option : All elements analyzed (Normalised)
Number of iterations = 4

Standard :

C CaCO₃ 1-Jun-1999 12:00 AM
O SiO₂ 1-Jun-1999 12:00 AM
P GaP 1-Jun-1999 12:00 AM
S FeS₂ 1-Jun-1999 12:00 AM
Ca Wollastonite 1-Jun-1999 12:00 AM
Cr Cr 1-Jun-1999 12:00 AM
Fe Fe 1-Jun-1999 12:00 AM
Ni Ni 1-Jun-1999 12:00 AM
Zn Zn 1-Jun-1999 12:00 AM

CA - CALCIUM
P - PHOSPHOROUS
ZN - ZINC
FE - IRON
CR - CHROMIUM
NI - NICKEL
C - CARBON
S - SULFUR
O - OXYGEN

Element	Weight%	Atomic%
CK	9.90	17.26
OK	44.43	58.11
PK	4.34	2.93
SK	10.69	6.98
Ca K	23.47	12.25
Cr K	0.85	0.34
Fe K	1.39	0.52
Ni K	0.89	0.32
Zn K	4.03	1.29
Totals	100.00	

Project Notes:

DEBRIS OFF EXHAUST
VALVE THAT WAS SHIP
TO CHRIS